



www.taidoc.com

泰博科技股份有限公司  
TaiDoc Technology Corp.

新北市24888五股區五工二路127號617樓  
E1 7E, No.127, Wugong 2nd Rd., Wugu Dist.,  
New Taipei City 24888, Taiwan

Tel : +886-2-6625-8185  
Fax : +886-2-6625-0288

## EC Declaration of Conformity

We, TaiDoc Technology Corporation

B1-7F, No.127, Wugong 2nd Road, Wugu Dist., 24888 New Taipei City, TAIWAN

declare under our sole responsibility that the product:

**Product Name** : Total Cholesterol Control Solution  
**Product Model** : TD-4919  
**Classification** : 98/79/EC (IVDD), Annex II, Self-testing  
**Conformity Assessment Route** : 98/79/EC (IVD), Annex IV excluding section 4 & 6  
**EC Certificate Number** : V1 052126 0069 Rev.03  
**European Representative** : MedNet EC-REP GmbH  
 Borkstraße 10, 48163 Münster, Germany  
**Notified Body (CE 0123)** : TÜV SÜD Product Service GmbH  
 Ridlerstraße 65, 80339 München, Germany  
**GMDN code** : 44697

to which this declaration relates is in conformity with the following standard(s) or other normative document(s):

EN ISO 13485:2016	Medical devices - Quality management systems - Requirements for regulatory purposes.
EN ISO 14971:2012	Medical devices - Application of risk management to medical devices
EN ISO 15223-1:2021	Medical devices - Symbols to be used with information to be supplied by the manufacturer - Part 1: General requirements
EN ISO 17511:2021	In vitro diagnostic medical devices - Requirements for establishing metrological traceability of values assigned to calibrators, trueness control materials and human samples.
EN ISO 18113-1:2011	In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 1: Terms, definitions and general requirements
EN ISO 18113-2:2011	In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 2: In vitro diagnostic reagents for professional use
EN ISO 18113-4:2011	In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 4: In vitro diagnostic reagents for self-testing
EN ISO 23640:2015	In vitro diagnostic medical devices. Evaluation of stability of in vitro



泰博科技股份有限公司  
TaiDoc Technology Corp.

新北市24888五股區五工二路127號01 7樓  
01 7F., No. 127, Wugong 2nd Rd., Wugu Dist.,  
New Taipei City 24888, Taiwan

Tel : +886-2-6625-8188  
Fax : +886-2-6625-0288

www.taidoc.com

	diagnostic reagents
EN ISO 20417:2021	Medical devices—Information to be supplied by the manufacturer
EN 13532:2002	General requirements for in vitro diagnostic medical devices for self-testing
EN 13612:2002+AC:2002	Performance evaluation of in vitro diagnostic medical devices

2022.5.6.

Date of Issue

**Jim Jan**  
Management Representative