

**DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC
CONCERNING MEDICAL DEVICES**

CONTEC MEDICAL SYSTEMS CO., LTD

MANUFACTURER: No.112 Qinhuang West Street, Economic & Technical Development Zone, Qinhuangdao, Hebei Province, PEOPLE'S REPUBLIC OF CHINA

MEDICAL DEVICE: Pulse Oximeter Probe, ESB0059

CLASSIFICATION - ANNEX IX: Class II b, Rule 10

CONFORMITY ASSESSMENT ROUTE: Annex II excluding chapter 4

WE, (CONTEC MEDICAL SYSTEMS CO., LTD) HEREWITHE DECLARE THAT THE STATED MEDICAL DEVICES MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES; INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EC. ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE. THIS EU DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE RESPONSIBILITY OF THE MANUFACTURER.

STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.

NOTIFIED BODY: TÜV SÜD PRODUCT SERVICE GMBH
RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY

IDENTIFICATION NUMBER:  0123

(EC) CERTIFICATE(S): G1 050972 0050 Rev.04

EUROPEAN REPRESENTATIVE: Prolinx GmbH
Brehmstr. 56, 40239, Duesseldorf, Germany

PLACE, DATE OF DECLARATION: QINHUANGDAO, 2024/04/01

SIGNATURE:



President

DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC
CONCERNING MEDICAL DEVICES

Appendix: list of (harmonised - EN) standards

No.	Standards	Title and Description
1	ISO 13485:2016	Medical devices Quality management systems Requirements for regulatory purposes
2	ISO 14971:2019	Medical devices Application of risk management to medical devices
3	EN60601-1:1990+A1:1993+A2:1995 (IEC60601-1:1988+A1:1991+A2:1995)	Medical electrical equipment- Part 1: General requirements for safety
4	EN 60601-1-2:2007 (IEC60601-1-2:2007)	Medical electrical equipment- Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests
5	EN 60601-1-6:2007 (IEC 60601-1-6:2006)	Medical electrical equipment-Part 1-6:General requirements for basic safety and essential performance-Collateral Standard: Usability
6	ISO10993-1: 2009	Biological evaluation of medical devices - Part 1: Evaluation and testing
7	EN ISO 9919:2009	Medical electrical equipment - Particular requirements for the basic safety and essential performance of pulse oximeter equipment for medical use
8	ISO 15223-1:2021	Medical devices Symbols to be used with information to be supplied by the manufacturer Part 1: General requirements
9	ISO 20417:2021	Medical devices Information to be supplied by the manufacturer