

## Declaration of Conformity



**Manufacturer:** Shenzhen Mindray Bio-Medical Electronics Co., Ltd.  
Mindray Building, Keji 12th Road South, High-tech Industrial Park,  
Nanshan, Shenzhen, 518057, P. R. China

**Manufacturer SRN:** CN-MF-000014156

**EC-Representative:** Shanghai International Holding Corp. GmbH (Europe)  
Eiffestraße 80  
20537 Hamburg, Germany

**Product:** Digital Ultrasonic Diagnostic Imaging System

**Model:** DP-10, DP-10T, DP-11, DP-15, DP-18, DP-20, DP-20T, DP-21, DP-25,  
DP-28, DP-30, DP-30T, DP-36

**Basic UDI-DI:** 69449040AB050100135N

**Classification:** IIa (According to Rule 10 of MDR Annex VIII)

**Conformity Assessment Route:** Annex IX excluding CHAPTER II

**CND code:** Z110401

**Supplementary information:** Included are following transducers: 35C20EA, 35C50EA, 65EC10EA,  
75L38EA, 65C15EA, 10L24EA, 35C50EB, 65EC10EB, 75L38EB,  
75L53EA.

**We declare that the above mentioned products meet the provisions of the REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT. All supporting documentations are retained under the premises of the manufacturer. This declaration of conformity is issued under the sole responsibility of the manufacturer.**

**References to CS:** /

**Notified Body:** TÜV SÜD Product Service GmbH  
Ridlerstraße 65  
80339 München, Germany.

**Notified Body No. :** 0123

**Identification of the Certificate:** G10 044751 0176

**Start of CE-Marking:** 2012-02-21

**I hereby am appointed as the authorized person to deal with all the registration and quality management affairs in my capacity as Deputy Director of Technical Regulation Department of Shenzhen**

**Mindray Bio-Medical Electronics Co., Ltd, Effective immediately.**

**Place, Date of Issue:** Shenzhen, 2023-3-31

**Signature:**

**Name of Authorized Signatory:** Mr. Wang Xinbing

**Position Held in Company:** Deputy Director, Technical Regulation

## Applied Standards List

**Product:** Digital Ultrasonic Diagnostic Imaging System

**Model:** DP-10, DP-10T, DP-11, DP-15, DP-18, DP-20, DP-20T, DP-21, DP-25, DP-28, DP-30, DP-30T, DP-36

### Standards Applied:

<b>EN ISO 14971:2019/A11:2021</b>	Medical devices – Application of risk management to medical devices
<b>EN ISO 20417 :2021</b>	Information supplied by the manufacturer of medical devices
<b>EN ISO 15223-1:2021</b>	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied – Part1: General requirements
<b>EN ISO 10993-1:2020</b>	Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process
<b>EN60601-1:2006/A1:2013</b>	Medical electrical equipment -- Part 1: General requirements for basic safety and essential performance
<b>EN60601-1-2:2015</b>	Medical electrical equipment -- Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests
<b>EN 60601-1-6: 2010/A1:2015</b>	Medical electrical equipment - Part 1-6: General Requirements for basic safety and essential performance -Collateral standard: usability
<b>EN 60601-2-37:2008/A1:2015</b>	Medical electrical equipment -- Part 2-37: Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment
<b>EN 62304:2006/A1:2015</b>	Medical device software - Software life-cycle processes
<b>EN 62366-1:2015</b>	Medical devices -- Application of usability engineering to medical devices
<b>EN ISO 17664-1:2021</b>	Processing of health care products - Information to be provided by the medical device manufacturer for the processing of medical devices
<b>ISO 17664-2:2021</b>	Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices — Part 2: Non-critical medical devices