DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

CONTEC MEDICAL SYSTEMS CO., LTD

MANUFACTURER: No.112 Qinhuang West Street, Economic & Technical

Development Zone, Qinhuangdao, Hebei Province,

PEOPLE'S REPUBLIC OF CHINA

MESH NEBULIZER, NE-M01

CLASSIFICATION - ANNEX IX: Class II a, Rule 11

CONFORMITY ASSESSMENT ROUTE: Annex II excluding chapter 4

WE, (CONTEC MEDICAL SYSTEMS CO., LTD) HEREWITH DECLARE THAT THE STATED MEDICAL DEVICES MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES;

INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EC. ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE.

THIS EU DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE RESPONSIBILITY OF THE MANUFACTURER.

STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.

NOTIFIED BODY:

TÜV SÜD PRODUCT SERVICE GMBH

RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY

IDENTIFICATION NUMBER: (€ 0123

(EC) CERTIFICATE(S): G1 050972 0050 Rev.04

Prolinx GmbH

EUROPEAN REPRESENTATIVE: Brehmstr. 56, 40239, Duesseldorf, Germany

PLACE, DATE OF DECLARATION: QINHUANGDAO, 2024-06-07

SIGNATURE:

President

TF-CE180403-09	Ver: D
Page 1 of 3	

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

Appendix: list of (harmonised - EN) standards

No.	Standards	Title and Description
1	EN ISO 13485:2016 (ISO 13485:2016)	Medical devices - Quality management systems - Requirements for regulatory purposes
2	ISO 14971:2019	Medical devices - Application of risk management to medical devices
3	IEC 60601-1:2005+ AMD1:2012+AMD2:2020	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance
4	IEC 60601-1-2:2014+AMD1:2020	Medical electrical equipment- Part 1-2: General requirements for safety and essential performance - Collateral standard: Electromagnetic disturbances - Requirements and tests
5	IEC 60601-1-11:2015 +AMD1:2020	Medical electrical equipmentPart 1-11: General requirements for basic safety and essential performance - Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment
6	EN13544-1:2007+A1:2009	Respiratory therapy equipment Part 1:Nebulizing system and their components
7	IEC 60601-1-6:2010+AMD1:2013 +AMD2:2020	Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral Standard: Usability
8	IEC 62366-1:2015+AMD1:2020	Medical devices - Application of usability engineering to medical devices
9	IEC 62304:2006+AMD1:2015	Medical device software-Software life-cycle processes
10	EN ISO15223-1:2021	Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements
11	ISO10993-1: 2018	Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process
12	ISO 18562-1:2017	Biocompatibility evaluation of breathing gas pathways in healthcare applications — Part 1: Evaluation and testing within a risk management process
13	ISO 18562-2:2017	Biocompatibility evaluation of breathing gas pathways in healthcare applications — Part 2: Tests for emissions of particulate matter
14	ISO 18562-3:2017	Biocompatibility evaluation of breathing gas pathways in healthcare applications — Part 3: Tests for emissions of volatile organic compounds (VOCs)
15	ISO 18562-4:2017	Biocompatibility evaluation of breathing gas pathways in healthcare applications — Part 4: Tests for leachables in condensate
16	ISO 10993-17:2002	Biological evaluation of medical devices — Part 17: Establishment of allowable limits for leachable substances

TF-CE180403-09	Ver: D
Page 2 of 3	

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

	17	ISO20417: 2021	Medical devices — Information to be supplied by the manufacturer	
--	----	----------------	--	--

TF-CE180403-09 Ver: D
Page 3 of 3