GIMA S.p.A. Via Marconi, 1 20060 Gessate (MI) –Italy www.gimaitaly.com



ITALIAN DIVISION gima@gimaitaly.com
EXPORT DIVISION export@gimaitaly.com

EU DECLARATION OF CONFORMITY

We, undersigned GIMA S.p.A. (single registration number (SRN): IT-MF-000011004), with operational headquarters in Gessate (MI), Via Marconi 1, and registered office in Milan, Via Tommaso Grossi 2, acting as manufacturer of the medical device:

| Product and trade name | Product code | Basic UDI-DI |
|---|--------------|---------------------------|
| PATIENT TROLLEY | 27800 | 80232790000V089906000008V |
| PATIENT TROLLEY - REMOVABLE TOP | 27801 | 80232790000V089906000008V |
| SIDE RAILS - pair for code 27800/1, 27850 | 27810 | 802327900Y1812270000000H9 |
| BASKET for code 27800 | 27812 | 80232790000L90990000010UC |
| OXYGEN BOTTLE HOLDER for code 27800/1 | 27813 | 80232790000V089900000006F |
| TRENDELENBURG TROLLEY | 27850 | 80232790000V089906000008V |
| OXYGEN BOTTLE HOLDER for code 27850 | 27854 | 80232790000V089900000006F |
| MATTRESS for code 27850 | 27855 | 802327900Y0333068100000FN |
| WARD STRETCHER without accessories | 44760 | 80232790000V089906000008V |
| SIDE RAILS - couple with automatic locking device | 44761 | 802327900Y1812270000000H9 |
| BASKET - chrome plated steel | 44763 | 80232790000L90990000010UC |
| OXYGEN BOTTLE HOLDER - chrome plated steel | 44764 | 80232790000V089900000006F |
| FIREPROOF SIMULATED LEATHER MATTRESS -class I | 44765 | 802327900Y0333060000000AZ |

Intended purpose:

Stretchers and related accessories: facilitate the transport of a patient lying down within a healthcare facility. The device can also be used as a treatment table or bed in some intensive care and specialist units or in conditions of high patient influx Side rails: reduce the risk of accidental patient falls from the bed

risk class I (not sterile), in accordance with the rule 1 set out in Annex VIII of the Regulation (EU) 2017/745 (MDR), declares, under its sole responsibility, that this device:

- complies with the Regulation (EU) 2017/745 (MDR);
- Common Specifications have not been used for the compliance of the above medical device.

Gessate, 27/06/2025

GIMA S.p.A.
The legal Representative
(Nicola Manzoni)