DC 0055 Page 1/2

OperSurgical

Declaration of Conformity

Application of Council Directive(s):	Medical Device Directive 93/42/EEC
Standards to which Conformity is Declared:	BS EN ISO 13485, BS EN 10993-1, BS EN 14971, BS EN 1041, BS EN ISO 15223-1
Manufacturer's Name:	CooperSurgical, Inc. also trading as Wallach Surgical Devices
Manufacturer's Address:	95 Corporate Drive, Trumbull, CT 06611 USA
Authorized European Representative:	EMERGO EUROPE Prinsessegracht 20 2514 AP The Hague The Netherlands
Type of Equipment:	Cervical Cell Collector
Device classification per MDD 93/42/EEC:	Class I Non-Sterile Rule 5
Conformity Assessment:	Annex VII of MDD 93/42/EEC
Model Numbers:	908001, 908002, 908003, 908004, 908006 (Refer to page 2 for details)
Date of First Issue:	July 26, 2010

We hereby declare that the equipment specified above conforms to the above Directive(s) and Standard(s). We also declare conformity against the United Kingdom Statutory Instrument SI 1994/3017 as amended by SI 2002/618.

Rev. B Aug 24, 2011 Removed references to notified body and conformity assessment; changed declaration language from individual to company ("I" to "We"); added document ID number and revise AER to include name Rev. C: Mar 1, 2012 Add Notified Body Rev. D: March 01, 2012 Add Conformity Assessment Annex VII Rev. E: July 01, 2013 Change BSI address Rev. F: Oct. 07, 2014 Updated Standards List, Manufacturer name and EU Rep Rev. G: Update Emergo address, expand Model Numbers and Device Description on Page 2, update J.C. Jensen's title to

(Signature) (Date) J. Chris Jensen (Full Name)

Sr. Director, Regulatory Affairs & Compliance (Position)

Jon. 10, 2018

Senior status – 06/20/2017 Rev. H: Update product descriptions on page 2.

Correct/Specify quantity description – 01/09/2018

95 Corporate Drive, Trumbull, CT 06611 USA Phone: (203) 799-2000 Fax: (203) 799-2002

Declaration of Conformity

Cervical Cell Collector

Model Number	Model Name
908001	Cervical Cell Sampler – QTY 100 per Bag
908002	Papette – QTY 5 per Bag
908003	Cervical Cell Sampler – QTY 1,000 per Box, 10 Bags of 100
908004	Cervical Cell Sampler – QTY 4,000 per Box, 40 Bags of 100
908006	Papette – QTY 500 per box, 20 Bags of 25

95 Corporate Drive, Trumbull, CT 06611 USA Phone: (203) 799-2000 Fax: (203) 799-2002