



## EUROPEAN MEDICAL DEVICE REGULATION

### Declaration of Conformity

As Legal Manufacturer, we

3M Deutschland GmbH  
Health Care Business  
Single Registration Number: DE-MF-000011641  
Carl-Schurz-Str. 1  
41453 Neuss  
Germany

hereby declare under our sole responsibility that the following CE marked device(s)

Trade Name	<b>Tegaderm™ + Pad</b>
Intended Purpose	Film dressing with non-adherent pad
Reference	<b>3582E, 3584E, 3586E, 3589E, 3590E, 3591E, 3593E, 3582NP, 3584NP, 3586NP, 3589NP, 3582SP, 3586SP, 3590SP, 3582P, 3586P, 3589P, 3590P, 3582P-10, 3586P-10, 3582IP, 3586IP, 3589IP</b>
Basic UDI-DI	06082232761010000000048D7

are classified per rule 4 of Annex VIII of the Medical Device Regulation (EU) 2017/745, as Class IIa sterile devices in accordance with Annex IX and all other applicable provisions of the REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL concerning medical devices.

This declaration is made based on the quality assurance certificate  
EU Quality Management Certificate: 003626MDR2017Q  
Issued by: DQS Medizinprodukte GmbH, No. 0297

February 21, 2023

Margaret Bessenbach  
Director Regulatory Affairs and Quality  
Health Care Business EMEA  
3M Deutschland GmbH

Date

3M is a trademark of 3M.

Related to REG-STED-MDR-05-855003