

**DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC
CONCERNING MEDICAL DEVICES**



MANUFACTURER:

CONTEC MEDICAL SYSTEMS CO., LTD

No.112 Qinhuang West Street, Economic & Technical
Development Zone, Qinhuangdao, Hebei Province,
PEOPLE'S REPUBLIC OF CHINA

MEDICAL DEVICE:

Pulse Oximeter Probe, ESA0015

CLASSIFICATION - ANNEX IX:

Class II b, Rule 1

CONFORMITY ASSESSMENT ROUTE: Annex II excluding chapter 4

WE, (CONTEC MEDICAL SYSTEMS CO., LTD) HEREWITH DECLARE THAT THE STATED
MEDICAL DEVICES. MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF
COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES;
INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC
ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE.

STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH
DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.

NOTIFIED BODY:

TÜV SÜD PRODUCT SERVICE GMBH
RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY

IDENTIFICATION NUMBER:

CE 0123

(EC) CERTIFICATE(s):

G1 050972 0050 Rev.04

EC REP

EUROPEAN REPRESENTATIVE:

Shanghai International Holding Corp. GmbH(Europe)
Eiffestrasse 80, 20537 Hamburg Germany

START OF CE-MARKING: 2009-07-23 (Date or Lot or serial number)

PLACE, DATE OF DECLARATION: QINHUANGDAO, 2020-06-18

SIGNATURE:

President