

DECLARATION OF CONFORMITY

We declare under own responsibility that the following scales:

- **Non-approved Medical Scales**

**7758, 7759, 7808, 7850,
2926, 9320, 7801, 7802, 7830, 7831, 7858, 7859, 7870, 7871, 7872, 8352**

- **Indicators**

2750, 3700

manufactured by us meet the provisions of the Council Directive

2014/30/EU

In general we meet the following directives:

Directive / Standard	
Company standard scale	FR 001 ²
EMV directive	2014/30/EU
Standard for emission	EN60601-1-2
Standard for immunity	EN60601-1-2
immunity criterion	A (3V/m) ⁴
ESD criterion	B
burst criterion	(B) ³
Surge criterion	(B) ³
Low voltage directive	2014/35/EU
Standard for electr. Safety	EN61010-1
Battery directive	2006/66/EC
REACH	(EG) No. 1907/2006
Food law	1935/2004/EC
Protection class	EN 60 529
RoHS	2011/65/EU

²⁾ FR 001 = Factory Regulation 001

³⁾ in the event that electrical wires are present external to the product


⁴⁾ according to article 4 of the EMC act, a comment in German must be contained in the operating instructions, if the criterion cannot be maintained

Please note:

This document annuls the previous sent Declaration of Conformity for non-approved medical scales and electronic height rods.

Validity: until 31.12.2017

**Soehnle Industrial Solutions GmbH
Gaildorfer Str. 6
D-71522 Backnang, Germany**



**Michael Schläger
General Manager
13.05.2016**