

EU DECLARATION OF CONFORMITY		
Product Group	Sterile Synthetic Absorbable Surgical Suture	
Generic Name	Braided and Coated Polyglactin 910 Fast	
EMDN Code	H0101010202	
Basic UDI-DI	8903837H0101010202ZK	
Brand Names	TRUSYNTH FAST, VETSUTURE FASTPGLA, NOÉCARE FASTPGLA, NOÉDENTAL FASTPGLA, ALPHA-CRYIL FAST, ISYNTH FAST, POLYGLACTIN FAST, Q-CLOSE SYNTH FAST, U-SYNTH FAST, SURGISUT TRUSYNTH FAST, TRUVET POLYGLACTIN FAST, FLYSORB FAST, IM-SYNTH FAST, ECRYL PLUS, UNODENT PGLA FAST, ALAN LACT FAST, POGAL 910 FAST, N-CARE POLYGLACTION 910 FAST	
Sizes	USP Sizes/EP Gauge - 8-0/0.4, 7-0/0.5, 6-0/0.7, 5-0/1, 4-0/1.5, 3-0/2, 2-0/3, 0/3.5, 1/4, 2/5	
only short term wound support is required and whe profile, Braided and Coated Polyglactin 910 fast abs	I fast absorbable suture is intended for use in general soft tissue approximation where re rapid absorption of the suture would be beneficial. Due to the rapid absorption orbable suture is useful for skin closure, particularly in paediatric surgery, sa. Braided and Coated Polyglactin 910 fast absorbable suture is also useful for	

Registered Place of Business:	Healthium Medtech Limited
	No. 472 D, 13 th Cross, 4 th Phase, Peenya Industrial Area,
	Bangalore – 560 058, Karnataka, India.
	SRN No.: IN-MF-000008421
Address of the Manufacturing Unit:	Site 01: Healthium Medtech Limited
	No. 472 D, 13th Cross, 4th Phase, Peenya Industrial Area,
	Bangalore Karnataka-560 058. India
	Site 02: Healthium Medtech Limited
	PLOT NO: 1605, PORTIA ROAD, SRI CITY, TIRUPATI DISTRICT,
	ANDHRA PRADESH– 517 646, INDIA
E.U. Representative's name and address	MED DEVICES LIFESCIENCES B.V.
	Keizersgracht 482
	1017 EG Amsterdam,
	The Netherlands
	Email:info@meddevices.net
	Phone: +31-202254558
	SRN No.:NL-AR-000000594
Classification of the device:	Braided and Coated Polyglactin 910 Fast sutures is classified as Class III
	device in accordance with Rule 8 of Annex VIII of the MDR 2017/745
Notified Body's Address and	3EC International a.s.
Notified body number	Hranicna 18, 821 05 Bratislava,
	Slovak Republic. NB 2265
Certificate No.:	2022-MDR/QS-016
Document Reference No.	TD/CE/04
Applicable Standards	EN ISO 13485:2016, ISO 14971:2019, ISO 9001: 2015,
- Ipp	EN 556-1: 2001, EN 556-2: 2015, EN ISO 15223-1: 2021, EN ISO 20417-
	2021, EN ISO 11135-2014, ISO 10993-1:2018, ISO-11607-1 : 2019, ISO-
	11607-2 : 2019, ISO-11737-1: 2018, ISO 11737-2: 2019, EN ISO
	14155:2020, ISO 14644-1:2015, EU.PHARMACOPEIA 10th version,
	MEDDEV 2.7-1 rev 4, MEDDEV 2.12-1 rev 8, MEDDEV 2.12-2 rev 2
Quality Paspansibility	

Quality Responsibility:

Healthium Medtech Limited, declares under sole responsibility for complying all the quality requirements of the product.

Declaration: Healthium Medtech Limited, declares that Braided and Coated Polyglactin 910 Fast sutures meets the requirements of Quality as per USP specifications given in the monograph, Absorbable Surgical Suture, Synthetic and also complies with MDR 2017/745

The EU declaration of conformity is issued under the sole responsibility of the manufacturer.

Healthium Medtech Limited

(Formerly known as Healthium Medtech Private Limited)



Signature:	Pour lawor Tower Town Medico	
Name and	Pankaj Dawar	
Position	PRRC	
Date	15-Nov-2022	
Place	Bangalore	

Healthium Medtech Limited

(Formerly known as Healthium Medtech Private Limited)