

EU DECLARATION OF CONFORMITY	
Product Group	Sterile Synthetic Absorbable Surgical Suture
Generic Name	Braided And Coated Polyglycolic Acid-Fast Absorbable
EMDN Code	H0101010201
Basic UDI-DI	8903837H0101010201ZH
Brand Names	TRUGLYDE FAST, LINX RPGA, POLYGLYCOLIC ACID FAST, U-GLYDE FAST, SURGISUT TRUGLYDE FAST, PGA RESOQUICK, TRUVET POLYGLYCOLIC ACID FAST, IM-GLYDE FAST, ESORB PLUS, UNODENT PGA FAST, POGAL FAST, N-CARE PGA FAST, ALAN GLYDE FAST
Sizes	USP Sizes/EP Gauge - 8-0/0.4, 7-0/0.5, 6-0/0.7, 5-0/1, 4-0/1.5, 3-0/2, 2-0/3, 0/3.5, 1/4, 2/5
Intended Use: Braided & Coated Polyglycolic acid fast absorbable suture is intended for use in general soft tissue approximation where only short term wound support is required and where rapid absorption of the suture would be beneficial. Due to the rapid absorption profile, Braided & Coated Polyglycolic acid fast absorbable suture is useful for skin closure, particularly in paediatric surgery, episiotomies, circumcision and closure of oral mucosa. Braided & Coated Polyglycolic acid fast absorbable suture is also useful for conjunctival suturing in ophthalmic surgery	
Registered Place of Business:	Healthium Medtech Limited No. 472 D, 13 th Cross, 4 th Phase, Peenya Industrial Area, Bangalore – 560 058, Karnataka, India. SRN No.: IN-MF-000008421
Address of the Manufacturing Unit :	Site 01: Healthium Medtech Limited No. 472 D, 13th Cross, 4th Phase, Peenya Industrial Area, Bangalore Karnataka-560 058. India Site 02: Healthium Medtech Limited PLOT NO: 1605, PORTIA ROAD, SRI CITY, TIRUPATI DISTRICT, ANDHRA PRADESH– 517 646, INDIA
E.U. Representative's name and address	MED DEVICES LIFESCIENCES B.V. Keizersgracht 482 1017 EG Amsterdam, The Netherlands Email:info@meddevices.net Phone: +31-202254558 SRN No.:NL-AR-000000594
Classification of the device:	Braided and Coated Polyglycolic acid sutures is classified as Class III device in accordance with Rule 8 of Annex VIII of the MDR 2017/745
Notified Body's Address and Notified body number	3EC International a.s. Hranicna 18, 821 05 Bratislava, Slovak Republic. NB 2265
Certificate No.:	2022-MDR/QS-015
Document Reference No.	TD/CE/02
Applicable Standards	EN ISO 13485:2016, ISO 14971:2019, ISO 9001: 2015, EN 556-1: 2001, EN 556-2: 2015, EN ISO 15223-1 : 2021 , EN ISO 20417-2021, EN ISO 11135-2014, ISO 10993-1:2018, ISO-11607-1 : 2019, ISO-11607-2 : 2019, ISO-11737-1: 2018, ISO 11737-2: 2019, EN ISO 14155:2020, ISO 14644-1:2015, EU.PHARMACOEIA 10th version, MEDDEV 2.7-1 rev 4, MEDDEV 2.12-1 rev 8, MEDDEV 2.12-2 rev 2.
Quality Responsibility: Healthium Medtech Limited, declares under sole responsibility for complying all the quality requirements of the product.	
Declaration: Healthium Medtech Limited, declares that Braided and Coated Polyglycolic acid fast absorbable sutures meets the requirements of Quality as per USP specifications given in the monograph, Absorbable Surgical Suture, Synthetic and also complies with MDR 2017/745. The EU declaration of conformity is issued under the sole responsibility of the manufacturer.	



Healthium Medtech Limited

(Formerly known as Healthium Medtech Private Limited)

Corporate Office: RMZ North Star, Cowrks, 12th Floor, Adjacent to RMZ Galleria Mall, Yelahanka, Bengaluru - 560 064, India

Registered Office: 472/D, 13th Cross, 4th Phase, Peenya Industrial Area, Bengaluru - 560 058, India

www.healthiummedtech.com | CIN : U03311KA1992PLC013831

Signature :	 
Name and Position	Pankaj Dawar PRRC
Date	15-Nov-2022
Place	Bangalore

Healthium Medtech Limited

(Formerly known as Healthium Medtech Private Limited)

Corporate Office: RMZ North Star, Cowrks, 12th Floor, Adjacent to RMZ Galleria Mall, Yelahanka, Bengaluru - 560 064, India

Registered Office: 472/D, 13th Cross, 4th Phase, Peenya Industrial Area, Bengaluru - 560 058, India

www.healthiummedtech.com | CIN : U03311KA1992PLC013831