



<b>CE DECLARATION OF CONFORMITY</b>
-------------------------------------

Fulfil Appendix VII of the rule 93/42/EWG concerning medical products.

We, company HERDEGEN SNC, 7 rue Eugène Freyssinet 77500 Chelles, FRANCE, certifies that the following products :

**MOBILITY :**

***AXILLARY CRUTCHES***

Luxury version, tall adult	Ref	220 002
Luxury version, adult	Ref	220 000
Luxury version, teenager	Ref	220 001
Asiacare, tall adult	Ref	220 802
Asiacare, adult	Ref	220 800
Asiacare, teenager	Ref	220 801

***ELBOW CRUTCHES***

***EVOLUTION***

Standard	Ref	200 0**
With large rubber tip PIVOFLEX	Ref	200 06*
With reflector	Ref	202 00*
With large rubber tip PIVOFLEX & reflector	Ref	202 06*

***ADVANCE***

Standard	Ref	200 9**
With large rubber tip PIVOFLEX	Ref	203 9**
With reflector	Ref	205 9**
With large rubber tip PIVOFLEX & reflector	Ref	206 9**

***PROGRESS I***

Standard	Ref	200 700
With large rubber tip PIVOFLEX	Ref	200 760

***PROGRESS II***

Standard	Ref	201 700
With large rubber tip PIVOFLEX	Ref	201 760

***TIKI***

Crutch for children or teenager	Ref	200 5**
---------------------------------	-----	---------

***TRIPOD & QUADRIPOD***

Tripod for adults	Ref	220 3**
Tripod for children	Ref	221 317
Quadripod for adults	Ref	220 4**
Quadripod for children	Ref	221 413

***BLIND STICKS***

Foldable in 4 parts, mailloche handle	Ref	250 900
Foldable in 3 parts, T handle	Ref	250 500
Curved	Ref	251 090



### **WALKING STICKS**

Anatomic walking stick	Ref	243 60*
Foldable & adjustable walking sticks	Ref	250 28*
T adjustable walking sticks, anodised grey	Ref	250 700
T adjustable walking sticks, painted	Ref	250 800
Walking sticks	Ref	100 *** or 103 ***
Stick with mailloche chromium handle	Ref	121 040

### **SEATING STICKS**

Seating stick in aluminium	Ref	116 001
Metal tripod 55 cm	Ref	115 550
Tripod wooden with leather seat with strap	Ref	115 450
Seat stick with leather	Ref	110 020
Quattro	Ref	116 80*

### **WALKING FRAMES**

Standard walking frame	Ref	221 ***
Reciprocal & foldable	Ref	220 ***
Foldable	Ref	223 ***
3 in 1	Ref	222 850
With 4 grips	Ref	223 840

### **ROLLATOR**

Rollator	Ref	243 8**
Foldable & adjustable	Ref	243 113

Fulfil the complete regulations of :

- The rule 93/42/EWG appendix VII and that these products belong to the class I.
- Directive 2007/47/EC.
- EN ISO 14971 edition 1 (2007) Medical devices – Application of risk management to medical devices.
- EN 12182 : Technical aids for disabled persons – General requirements and test methods

Stutzheim, 5<sup>th</sup> March 2012

Hervé MEYER

**HERDEGEN SNC**  
7 rue Eugène Freyssinet  
F - 77500 Chelles  
Tél : +33 (0)1 64 21 70 33 - Fax : +33 (0)1 64 21 71 08  
SIREN 50860164800019 - APE 3250A - TVA FR 00 238 601 648  
SNC au Capital de 10.000 Euros

A handwritten signature in black ink, appearing to read 'H. MEYER', is placed to the right of the company contact information.