



of Conformity

Registration No.:

AK 50144621 0001

Report No.:

12307539 001

Holder:

RF Co., Ltd. 3 Nakagosho Nagano-shi, Nagano 380-0935 JAPAN

Product:

X-Ray Equipment

Multi CCD Digital Radiography NAOMI

Identification:

Type Designation: NAOMI MID

Serial Number : MD046036H

Tested acc. to:

EN 60601-1-2:2001+A1 IEC 60601-1-2:2005

The certificate of conformity refers to the above mentioned product. This is to certify that the specimen is in conformity with the assessment requirement mentioned above. This certificate does not imply assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity.

Yokohama, 07.01.2009

TÜVRheinland fication

Certification Body

Dipl.-Ing. M. Geiser

TÜV Rheinland Japan Ltd. - Yokohama 222-0033 - JAPAN