

## EC Declaration of Conformity

**Manufacturer:**

**Name:** Xianning Full Guard Medical Products Co., Ltd

**Address:** Yongan East Avenue, Xianning Economic Development Zone, Xianning City, Hubei Province, China

**Tel/Fax:** 0715-8200113

**SRN:** Not available yet

**Basic -UDI -DI** 697386655FGIS-PP-LXZ

**Whose single Authorized Representative:**

**Name:** ZOUSTECH S.L

**Address:** Pso.Castellana, 141- planta 19, 28046-Madrid, Spain

**Tel/Fax:** +34694426446

**SRN:** Not available yet

**Disposable Isolation Gown**

UMDNS CODE: 15037

Product Code:FGIS-PP25-M,FGIS-PP25-L,FGIS-PP25-XL, FGIS-PP25-XXL

Product Size: M,L,XL, XXL

Classification According To MDR, Annex VIII: Class I – Rule 1

Conformity Assessment Route: Annex II and Annex III

Applied Common Specification/Standard:

EN ISO 14971:2019 Medical devices - Application of risk management to medical devices (ISO 14971:2019)

ISO 10993-1:2018 Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process



EN ISO 10993-5:2009 Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity (ISO 10993-5:2009)

EN ISO 10993-10:2013 Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization (ISO 10993-10:2010)

EN ISO 15223-1:2016 Medical devices—Symbols to be used with medical device labels, labeling and information to be supplied.

We, the manufacturer, herewith declare under our sole responsibility that the above mentioned product, meets the provision of Regulation (EU) 2017/745 on Medical Devices (MDR). All supporting documentations are retained under the premises of the manufacturer.


### REGULATION

REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 5 April 2017 on medical devices.



Place of Issue: Xianning, Hubei 湖北咸宁

Date of Issue: 2021.4.21

Signature: 

Name: Rosen Jiang

Position: Managing Director

Stamp



Issued By: Molly Zheng

Approved By: Rosen Jiang

Revision: FD01

Effective Date: 2021.4.21