

Declaration of Conformity

Date: 13 November 2013

Manufacturer: Nonin Medical, Inc.
Address: 13700 1st Avenue North
Plymouth, MN 55441-5443 USA

**Model Numbers and
Product Designations:** 2500A
PalmSat® Pulse Oximeter System

Device Category(ies): Oximeter, pulse

GMDN Number(s): 17148

Date Added: January 2005

We herewith declare that the above mentioned pulse oximeter system is classified as Class IIb (using rule 10) and complies with Annex II, section 3 of the Directive No. 93/42/EEC on medical devices, as amended (2007/47/EC) and all applicable clauses of ISO 9919:2005, *Medical electrical equipment -- Particular requirements for the basic safety and essential performance of pulse oximeter equipment for medical use.*

This declaration applies to all devices manufactured from the date of issuance until it is either superseded by another declaration or withdrawn.

Notified Body: TÜV SÜD Product Service GmbH
Ridlerstrasse 65
D-80339 München
Germany

EC Certificate Number: G1 13 09 24497 023

Signature: 
Name: Kim E. Aves
Title: Senior Regulatory Affairs Specialist

Authorized EC Representative:
Medical Product Service (MPS) GmbH
Borngasse 20
35619 Braunfels
Germany