

DECLARATION OF CONFORMITY
TO COUNCIL REGULATION (EU) 2017/745 CONCERNING
MEDICAL DEVICES

MANUFACTURER:

Shenzhen Creative Industry Co., Ltd.
1001, Building West, Lepu Tower, No.66 Xingke Road,
Xili Community, Xili Street, Nanshan District,
518055 Shenzhen, PEOPLE'S REPUBLIC OF CHINA

SRN OF the MANUFACTURER:

CN-MF-000009430

MEDICAL DEVICE:

SpO2 Probe Cable Extender

MODEL:

2909-0000004/2909-0000034/2909-0000035/2909-0000038
15040012

INTENDED USE:

The Cable Extender is used to connect a compatible
patient monitor or a pulse oximeter device and a
SpO2 Probe. It ensures that there is enough length for
the SpO2 Probe to acquire oxygen saturation(SpO2)
and pulse rate(PR) from a patient.

CLASSIFICATION - ANNEX VIII:

Class I, Rule 1

BASIC UDI-DI:

69419006CableExtender3U

CONFORMITY ASSESSMENT ROUTE:

Annex II+ Annex III

WE, **Shenzhen Creative Industry Co., Ltd.**, HEREWITH DECLARE THAT THE STATED MEDICAL DEVICES MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL REGULATION (EU) 2017/745 CONCERNING MEDICAL DEVICES; ALL SUPPORTING DOCUMENTATION ARE RETAINED UNDER THE PREMISES OF THE MANUFACTURER. WE ARE EXCLUSIVELY RESPONSIBLE FOR THE DECLARATION OF CONFORMITY.

STANDARDS APPLIED:		
EN ISO 15223-1:2021	EN ISO 20417:2021	EN ISO 10993-1:2020
EN ISO 10993-5:2009	EN ISO 10993-10:2013	ISO 10993-23:2021

EC

REP

EUROPEAN REPRESENTATIVE:

Shanghai International Holding Corp. GmbH (Europe)
Eiffestraße 80, 20537 Hamburg, Germany


SRN OF the EUROPEAN REPRESENTATIVE:

DE-AR-000000001

PLACE, DATE OF DECLARATION:

Shenzhen, Oct.17, 2025

SIGNATURE:



NAME:

Zhang Xiang

POSITION:

Management Representative