


**DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC
CONCERNING MEDICAL DEVICES**

| | |
|-------------------------------------|--|
| MANUFACTURER: | CONTEC MEDICAL SYSTEMS CO., LTD. No.112 Qinhuang West Street, Economic & Technical Development Zone, Qinhuangdao, Hebei Province, PEOPLE'S REPUBLIC OF CHINA |
| MEDICAL DEVICE: | Electrocardiograph, ECG300G |
| CLASSIFICATION - ANNEX IX: | Class II a, Rule 10 |
| CONFORMITY ASSESSMENT ROUTE: | Annex II excluding chapter 4 |

WE, (CONTEC MEDICAL SYSTEMS CO., LTD) HERewith DECLARE THAT THE STATED MEDICAL DEVICES MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES; INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EC. ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE. THIS EU DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE RESPONSIBILITY OF THE MANUFACTURER.

STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.

| | |
|---------------------------------|--|
| NOTIFIED BODY: | TÜV SÜD PRODUCT SERVICE GMBH RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY |
| IDENTIFICATION NUMBER: | CE 0123 |
| (EC) CERTIFICATE(S): | <u>G1 050972 0050 Rev.04</u> |
| EUROPEAN REPRESENTATIVE: | Prolinx GmbH Brehmstr. 56, 40239, Duesseldorf, Germany |

| | |
|------------------------------------|---|
| PLACE, DATE OF DECLARATION: | QINHUANGDAO, 2024/01/10 |
| SIGNATURE: |  _____ President |

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

Appendix: list of (harmonised - EN) standards

| No. | Standards | Title and Description |
|-----|--|---|
| 1 | IEC 60601-1:2005+ A1:2012+A2:2020 | Medical electrical equipment - Part 1: General requirements for basic safety and essential performance |
| 2 | IEC 60601-1-6:2010+ A1:2013+A2:2020 | Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral Standard: Usability |
| 3 | IEC 60601-2-25:2011 | Medical electrical equipment –Part 2-25: Particular requirements for the basic safety and essential performance of electrocardiographs |
| 4 | IEC 60601-1-2:2014+A1:2020 | Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests |
| 5 | IEC 62304:2006+A1:2015 | Medical device software-Software life-cycle processes |
| 6 | ISO 14971:2019 | Medical devices - Application of risk management to medical devices |
| 7 | EN ISO 15223-1:2021 | Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied -- Part 1: General requirements |
| 8 | ISO 20417:2021 | Medical devices - Information to be supplied by the manufacturer |
| 9 | IEC 62366-1:2015 +A1:2020 | Medical devices - Application of usability engineering to medical devices |