




PROFESSIONAL MEDICAL PRODUCTS

DIGITAL BLOOD PRESSURE MONITOR

INSTRUCTION MANUAL

Parts and Components



- 1. Main Body
- 2. Display
- 3. Air Connector
- 4. Tube plug
- 5. Air Hose
- 6. Cuff
- 7. D-ring
- 8. Button 'PRESET'
- 9. Button 'MEM1'
- 10. Button e'MEM2'
- 11. Button 'START'
- 12. Storage Case
- 13.4xAA Batteries(Optional)
- 14.AC Adapter(Optional)

SYMBOLS

Symbols	Meaning
	Manufacturer
	Authorized Representative in the European community
	WEEE disposal
	Medical Device complies with Directive 93/42/EEC
	Keep in a cool, dry place
	Follow instructions for use
	Type BF applied Part
	Stand by
	Lot number
	Product code
	Date of manufacture
	Importer

GENERAL

This instruction manual is intended to assist the user for safe and efficient operation of the automatic digital blood pressure monitor (hereinafter: device) model LD-582.The device must be used in accordance with the procedures described in the manual. It is important to read and understand the entire manual, especially the section <IMPORTANT SAFETY INSTRUCTIONS >.This device is intended for the non-invasive measurement of systolic and diastolic arterial blood pressure and pulse rate in adults (age 15 and above).

CAUTION:

- 1. Do not use this device on infants or persons who cannot express their intentions
- 2. The device is not suitable for measuring the blood pressure of children. Ask your doctor before using it on elder children.
- 3. The patient is an intended operator. But persons who suffer from arrhythmia,diabetes, cardiovascular problems or who have had a stroke should consult your doctor before using the device.

PRINCIPLE OF OPERATION

This device adopts the oscillometric technology with Fuzzy Algorithm to measure the arterial blood pressure and pulse rate. The cuff is wrapped around the arm and automatically inflated by the air pump. The sensor of the device catches weak fluctuation of the pressure in the cuff produced by extension and contraction of the artery of the arm in response to each heartbeat. The amplitude of the pressure waves is measured, converted in millimeters of the mercury column, and is displayed by digital value.

**ATTENTION:** This device can not provide reasonable accuracy if used or stored the temperature, humidity or altitude beyond the range stated in the section <SPECIFICATIONS> of this manual.

NEW TECHNOLOGIES USED

Fuzzy Algorithm is the processing algorithm, taking into account the specialty of individual heartbeats, which provides higher accuracy of measurement.

Software version:V1.1

IMPORTANT SAFETY INSTRUCTIONS

It is necessary to know that arterial blood pressure is subjected to sharp fluctuations. The level of the arterial blood pressure depends on many factors.Generally arterial blood pressure is lower in summer and higher in winter. Arterial blood pressure changes with atmospheric pressure and is affected considerably by many factors, e.g.physical loads, emotional excitability, stress, meals, etc.Medicines, drinking, smoking affect greatly the level of an individual's blood pressure.Blood pressure does vary with age and individuals, and it is recommended to write down the readings from blood pressure records daily, then you can check with your doctor to find out what is a "normal blood pressure measurement" for you. Please read the instruction manual carefully before using this device, especially <Important safety instructions>, it can help you use the device correctly and safely!Please keep the instruction manual for future use. For specific information about your own blood pressure, consult your physician.

WARNINGS

- Consult your physician if you suffer from illnesses prior to using the device.
- The device is not suitable for persons who have electrical implants.If you had a mastectomy (breast amputation) do not use this blood pressure monitor on the arm on the side of the mastectomy.
- Pregnant women should only measure their own blood pressure in consultation with their doctor, since the readings may be changed with pregnancy.
- Do not service or maintain the cuff while in use with patient.
- Do not use this blood pressure monitor on any arm where intravascular access or therapy (such as an intravenous drip or a blood transfusion), or an arteriovenous shunt (A-V shunt) is present. The temporary interference to blood flow by the blood pressure measurement could result in injury.
- Do not use the device with other medical electrical(ME) equipment simultaneously.
- Do not use the device in the area the HF surgical equipment, MRI, or CT scanner exists, or in the oxygen rich environment.
- Do not use a mobile phone or other devices that emit electromagnetic fields, near the device. This may result in incorrect operation of the device.
- Never use any accessories or parts from other manufacturers, Using such accessories or parts could cause a hazardous situation for the user or damage to the device.

- Do not modify this equipment without authorization of the manufacturer.
- The batteries used in this device may present a fire or chemical burn hazard if mistreated. Do not disassemble, heat or incinerate.
- Keep equipment away from fire and heat sources to prevent fire or explosion.
- Please keep the unit out of reach of infants, children or pets, since inhalation or swallowing of small parts can be dangerous or even fatal.
- Please pay attention that the continuous CUFF pressure due to connection tubing kinking will cause a harmful injury.
- Do not use an extension cord with this device.
- The air tube or the AC adapter cable may cause accidental strangulation in infants.
- Do not put the air tube around your neck- danger of suffocation!A device should never be left unattended when plugged in.
- Do not reach for a corded device that has fallen into water. Unplug immediately.
- It is quite normal that two measurements taken in quick succession may produce significantly different results, because too frequent and consecutive measurements could cause disturbances in blood circulation and injuries

Cautions

- Use this device under the right environmental conditions as indicated in this user manual. If not, this could affect the performance, lifetime of the device and measurement results.
- Only use this device for its intended purpose as described in this user manual.
- Do not confuse self-monitoring with self-diagnosis. This device allows you to monitor your blood pressure. Do not begin or end medical treatment based on the measurement results. Always consult your physician for treatment advice.
- Do not take any therapeutic measures on the basis of a self-measurement.Never change prescribed medication without consulting your physician. Consult our physician if you have any questions about your blood pressure.
- If you are taking medication, consult your physician to determine the most appropriate time to measure your blood pressure.
- Consult the physician if measurement errors occur in children or persons with arrhythmia.
- The pulse display is not suitable for monitoring the frequency of cardiac pacemakers.
- Common arrhythmias (such as atria or ventricular premature beats or atrial fibrillation) and peripheral artery disease /arteriosclerosis can affect the accuracy of this blood pressure monitor. Please consult your physician how to best use this blood pressure monitor if you suffer from any of these conditions. Blood pressure measurement is not suitable in cases of serious arteriosclerosis (hardening of the arteries).
- The effectiveness of this blood pressure monitor has not been established in pregnant women.
- Always check the device and cuff before you use it. Do not use the device or cuff if one of them is damaged, because this may cause injury.
- This device is not intended for use on extremities other than the arm or for functions other than obtaining a blood pressure measurement
- Do not attach the cuff on the same arm on which other monitoring medical electrical equipment is attached simultaneously, because this could cause temporary loss of function of those simultaneously-used monitoring medical electrical equipment.
- Never attach the cuff on injured skin, an injured arm or an arm under medical treatment as this can cause further injury.
- Do not forcibly crease the arm cuff or the air tube excessively.
- Do not press the air tube while taking a measurement.
- Do not use the device in case of existing polyester or nylon material allergies.
- This device is not suitable for continuous monitoring during medical emergencies or operations.
- This device cannot be used with HF (High Frequency) surgical equipment at the same time.
- This device is not washable. Never immerse the device in water and do not rinse it under the tap.
- This device should keep dry to prevent from moisture.
- The equipment is not AP/APG equipment and is not suitable for use in the presence of a flammable anesthetic mixture with air, with oxygen or nitrous.
- To avoid measurement errors, do not use the device near strong electromagnetic fields, radiated interference signal or electrical fast transient/burst signal. For example magnets, radio transmitters, microwave ovens.
- If this device was stored in low temperature, leave it in room temperature for at least 1 hour.
- Repeated measurements with an interval of 3 minutes are recommended, so you can calculate the average to get a more accurate measurement. An interval of 3minutes can also ensure that the operation of the device does not result in prolonged impairment of the circulation of the blood.
- Atherosclerosis patients may require longer interval (10-15minutes) as elasticity of patient's vessels decreases significantly with the disease.10-15minutes interval is also applicable for patients suffering from diabetes for a long period of time.
- Dispose of the device, components and optional accessories according to applicable local regulations. Unlawful disposal may cause environmental pollution.
- Connecting electrical equipment to mso effectively leads to creating a ME system,and can result in a reduced level of safety

CLASSIFICATION

- ME EQUIPMENT not intended for use in an oxygen rich environment or in the presence of flammable mixers.
- Internally powered equipment (without adaptor), Class II equipment (with adapter).
- Type BF applied part, recognize the cuff as applied part

BATTERY INSTALLATION

- 1. Open the battery cover and then open the button battery cover, then install one 'CR2025'button battery into the button battery compartment;
- 2. Close the button battery compartment cover;
- 3. Install four 'AA' type batteries into the battery compartment as indicated. Make sure that the polarity is correct;
- 4. Close the battery compartment cover.
- Inbuilt button battery for keeping the date/time uninterruptedly during changing the batteries(4x AA batteries).If the new batteries are installed into the device,the date and time displays '01/01' and '00:00' icon in the LCD, it indicates that you need change the new button battery.
- Replace the batteries when the replacement indication ' ' appears in the display or nothing after START button is pressed:
- Batteries in this kit are intended to check work capacity of the device and the life-span of the batteries can be shorter than the recommended
- Replace all batteries simultaneously, and don't use rechargeable batteries;
- Only same type batteries are allowed to be used together;
- If the device is to be unused for long time, please take out the batteries. Don't leave the worn batteries in the device;
- When the low battery indication' ' appears in the display or nothing after START button is pressed;
- Batteries in this kit are intended to check work capacity of the device and the life-span of the batteries can be shorter than the recommended;
- Replace all batteries simultaneously, and don't use rechargeable batteries;
- Only same type batteries are allowed to be used together;
- If the device is to be unused for long time, please take out the batteries.Don't leave the worn batteries in the device;
- When the low battery indication' ' flashes on the LCD during measurement, it reminds that the user will change all the batteries but can be used currently;;
- when the low battery indication' 'in the LCD and at the same time the buzzer beeps for 4 times continuously, it indicates that the user need change all batteries at once

USE AC POWER ADAPTER

Besides batteries you can use AC power adapter as the power supply. The AC power adapter is optional for this device for sale. The AC adapter is specified as a part of the blood pressure monitor.

- Insert the AC adapter cord into the jack on the right side of the monitor.
- Insert the AC adapter plug into the outlet.
- To remove the AC adapter, disconnect the adapter plug from the AC outlet first and then disconnect the cord from the monitor's jack.

Caution


- When using optional AC adapter, the AC adapter must comply with the requirements of standard IEC60601-1.
- To avoid possible damage of the monitor, use only the exclusive AC adapter that can be purchased from authorized dealers. Other adapter may damage the blood pressure monitor.
- The AC adapter is used as an isolating means, the AC adapter plug shall insert into the outlet nearby the operator, make it easy to disconnection the device from the outlet.
- If long time work, remove the plug after the adapter cools, and prevent burns.
- Plug the AC adapter into the appropriate voltage outlet. Do not use in a multi-outlet plug.
- Not to position the blood pressure monitor to make it difficult to operate the disconnection device( adaptor).

**Note: The monitor is designed not to draw power from the batteries when the AC adapter in use.**



Optional AC adapter technical feature:  
Model: YS5M-0600600  
Input: 100-240V 50/60Hz  
Output voltage: 6V+5%  
Output current: 600 mA  
Output plug polarity: <-> inner

ASSEMBLY THE STORAGE CASE



- 1. Three hooks of storage case aim at the concave of device respectively;
- 2. Push the storage case upwards;
- 3. To fill tightly with the plug.

SETTING DATE AND TIME

The function provides accurate measuring time for each measurement. To get accurate date and time, the user should preset the date and time correctly before the first use of this device. The operation procedure for presetting Date/Time is as follows:

- 1. When the device is connected to power supply at first time, the display will show as Fig.1:



Fig. 1



Fig. 2



Fig. 3

- 2. Press button 'PRESET', and the year number flashes;
- 3. Press button 'MEM1'or 'MEM2' to subtract or add the number, and press button 'START' for confirmation;
- 4. When the year setup is finished, the month number will flash automatically as Fig.3.Please follow the same instruction as above to set month,date and time;
- 5. Press button 'START' to finish setup. If you want to change the date and time,please repeat procedure 2, 3, 4.

FUNCTION OF REMINDERS

SETTING REMINDERS

This monitor has 3 reminder alarms. You can set 3 different reminder alarms within a 24 hours period.

- 1.When the device stands by, press button 'PRESET' two times to enter into alarm 01 mode, the display will show as Fig.4;

- 1. Press button 'MEM1'or'MEM2' the display will show as Fig.5 and at the same time the hour number flashes;
- 2. Press button 'MEM1'or'MEM2' again to subtract or add the number, and press button 'START' for confirmation.
- 3. When the hour number setup is finished, the minute number will flash automatically.Please follow the same instruction as above to set minute number;
- 4. Press button 'START' for confirmation.
- 6.When the device stands by, press button 'PRESET' three and four times respectively to enter into alarm 02 and 03 mode. Repeat the above.




Fig. 4

**Annotation: When the alarm is on under device stands by, the icon ' ' flashes on LCD and goes with beep for 1 minute,Press the button 'START' to turn off the alarm.**

**When the alarm is on during measurement, the icon' 'flashes on LCD for 1 minute without beep. Under this situation, if you press the button 'START',it will stop both the icon' ' flashing and measurement.**




Fig. 5



Fig. 6



Fig. 7




Fig. 7

REMINDERS CLEARANCE

- 1. When the device stands by, press button 'PRESET' two times to enter into alarm 01 mode, then press button 'MEM1' for at least 5 seconds, the display will show as Fig.7 which means the alarm 01 is removed.
- 2. When the device stands by, press button 'PRESET' three and four times respectively to enter into alarm 02 and 03 mode . Repeat the above process to remove the alarm 02 and alarm 03.

AMBIENT TEMPERATURE DISPLAY AND ADJUSTMENT

This monitor can display the ambient temperature and the unit °C and °F can be adjustable. °C mode display in the LCD when the first time used.

- 1. When the device stands by, press button 'PRESET' five times to enter into the temperature adjustable mode , then press button 'MEM1 to turn into F mode and press button 'START' for confirmation.
- 2. Press button 'MEM2 to convert °F mode into °C mode.

**Annotation: When under the mode of function reset,if without any operation in 1 minute, the device will automatically return standby mode.**




Fig.8


**Please keep quiet for 5-10 minutes and avoid eating, drinking, alcohol, smoking,exercising and bathing before taking a measurement, All these factors will influence the measurement result.**

- Remove any garment that fits closely to your upper arm.
- Always measure on the same arm(normally left).
- Measurements should be taken regularly at the same time of each day, as the blood pressure varies even during the day.
- Any effort to support the arm during measurement may increase the measured blood pressure.
- Make sure, you are in a comfortable, relaxed position with leg-uncrossed, feet flat on the floor, back and arm supported, middle of the cuff at the level of the right atrium of the heart and do not move or constrict your muscles and talk during measurement, Use a cushion to support your arm if necessary, Keep position in normal use.
- If the arm artery lies lower or higher than the heart, a false reading will be obtained.
- A loose or open cuff causes false readings.
- With repeated measurements, blood accumulates in the arm which can lead to false reading.
- Consecutive blood pressure measurements should be repeated after 1 minute pause or after the arm has been held up in order to allow the accumulated blood to flow away



CORRECT POSTURE

- 1. Sit beside the table and let the table support your arm as you take the measurement.
- 2. Sit upright with your back straight.
- 3. Make sure that the cuff on the upper arm no cross, and is at approximately the same level as the heart.
- 4.Make sure that your feet lie on the ground and no cross.
- 5.You may lie on your back and take a measurement.Look at the ceiling, keep calm, and don't move your neck or body during the measurement.



ASSEMBLY THE CUFF

- 1. Insert the edge of the cuff approximately 5 centimeters into the D-ring as shown in figure.
- 2.Put the cuff on the left upper arm with the tube pointing to the direction of palm.If measurement on your left arm is difficult, you can use right arm for measurement. In this case, it is necessary to know that the readings may differ about 5-10 mmHg between left arm and right arm.
- 3. Wrap cuff around your upper arm with the lower edge of the cuff approximately 2-3 centimeters above the elbow. The mark <ARTERY> must be over the artery of the arm.
- 4. Press the cuff to make sure that it is attached securely. The cuff should not be too tight or loose is greatly recommended. Two fingers should be easily put in between cuff and upper arm.
- 5. The mark <INDEX> on the cuff must point to area<NORMAL>Or <LARGE CUFF>.This means the cuff size is correct, If mark <NDEX> points to the area beyond area <NORMAL>or <LARGE CUFF>. please consult your dealer whether you need another size cuff. This device is supplied with the standard cuff which is fit for the arm size 22-32 cm.
- 6.Sometimes it is difficult to make the cuff regular depending on the shape of the user's upper arm. The cone-shape assembly of cuff is also acceptable.
- 7.If your clothes restrict the blood circulation of your upper arm, or you roll your sleeve up so as to result in such restriction,Please take off your shirt to get an accurate measurement if necessary.



Caution:

If you experience discomfort during a measurement such as pain in the upper arm or other complaints, press the 'START' button to release the air immediately from the cuff. Loosen the cuff and remove it from your arm.

CARRYING OUT A MEASUREMENT

- 1. Insert the tube plug into the air connector.Before the measurement, take 3-5 times deep breath and relax yourself. Don't talk or move your arm;
- 2. Press button 'START',and all symbols will appear on display in 2seconds as Fig.9. Then two short beep will sound and '0' will appear on the screen, Pump begins to inflate with display showing the reading of pressure. Generally the pressure will reach 190mmHg as Fig.10;
- 3. The pump stop inflating and pressure begins to decrease gradually during which the user's blood pressure and pulse will be calculated as Fig.11;



Fig. 9



Fig.10



Fig.11

- 4. There will be along beep following the accomplishment of measurement, The air in the cuff will deflate quickly and the blood pressure reading, pulse reading will show in the display. Moreover the measuring time will also display together in two screens alternately. At the same time,the '👤' will flash to remind the user to record the reading as Fig.12;
- 5. Press button'MEM1'or button'MEM2'to record the reading in corresponding memory, For example, if button 'MEM2'is pressed, the display will show as Fig.13. If the user does not press button, the reading won't be recorded;




Fig.12






Fig.13

- 6.Press the button 'START' to return to standby mode. Please rest for at least 3 minutes for another measurement. The device keeps unused for3 minutes, the device will be return to standby mode automatically.
- 7. If irregular heartbeat was detected during the measurement, LCD display the icon '♥' to remind users of heartbeat irregularity.
- NOTE:** The device will inflate to a higher pressure automatically in case the inflation pressure is not sufficient to determine measurement result.
- Attention: We recommend contacting your physician if you see the '♥' indicator frequently.

WHO CLASSIFICATION INDICATION

Standards for assessment of high or low blood pressure, regardless of age, have been established by World Health Organization(WHO) as show in the chart as below



The indicator display a segment, based on the current data, corresponding to the WHO classification. For example, if your blood press is 145mmHg (systolic pressure), 88mmHg(diastolic pressure), according to the world health organization standard, your blood pressure level is Mild Hypertension. Note:If the systolic blood pressure and the diastolic blood pressure fall into different categories, the higher value should be taken for classification.

FUNCTION OF MEMORY

MEMORY RECALL

- 1. LD-582 can store 60 sets of readings each in '👤' and '👤', and will automatically calculate the average value of the latest 3 readings for MEM1'and 'MEM2' respectively. When the memory is full (60 sets of readings are stored), the oldest reading will be replaced by new one. Memory will not clear away even if power supply is removed;
- 2. After a measurement is finished or when the device stands by,the user can press button 'MEM1'or button 'MEM2' to recall memory.Press button'MEM1' or 'MEM2', the display will show the average value of the latest 3readings as Fig.14;



Fig.14



Fig.15




Fig.16

- 3. Press again,the display will show '01', which means the latest reading, then turns to another screen to show readings and measuring time as Fig.15;
- 4.Press again, the display will show '02', which means the second to the latest reading...

MEMORY CLEARANCE

After a measurement is finished or when the device stands by, hold down button'MEM1'or'MEM2' for at least 5 seconds,the display will show 'CLR' which means the stored reading for'MEM1'or'MEM2' is removed.



ERROR AND LOW BATTERYINFORMATION

INDICATION	POSSIBLE REASON	CORRECTION METHODS
	The cuff is put on wrongly or the tube plug is inserted too loosely. Movement of arm/hand or talking during measurement. The cuff is not inflated to necessary pressure.	Make sure that cuff is put on correctly and the tube plug is inserted tightly and repeat the measurement.Repeat the measurement with following completely recommendations of manual. Repeat the measurement with pumping cuff to higher pressure.
	The batteries are weak	Replace all 4 batteries with new ones.

TROUBLESHOOTING

SYMPTOM	CHECK POINT	REMEDY
No display when connect the power	The batteries have run down.The polarity of battery is wrong.The contact of battery compartment is polluted.	Replace all the batteries with new ones.Install the batteries correctly. Clean the battery terminals with dry cloth.
Inflation stops and reinflate later.	The automatic inflation for ensuring correct measurement.Did you talk or move your arm(or hand) during measurement?	See<AUTOMATICINFLATION>Keep quiet and silent during the measurement.
The readings extremely low or high.	Is the cuff at the same level as the heart?Is the cuff wrapped right?Did you strain your arm during measurement?Did you talk or move your arm(or hand) during measurement?	Make sure that your posture is right. Wrap the cuff correctly.Relax during measurement.Keep quiet and silent during the measurement.
Pulse rate is too low or too high.	Did you talk or move your arm(or hand) during measurement?Did you make measurement right after exercise?	Keep quiet and silent during the measurement.Take measurement again after resting for more than 5minutes.
The batteries are run down soon.	Faulty batteries are used.	Use alkaline batteries of known manufacturers.

CARE,STORING,REPAIR AND RECYCLING

- 1. It's necessary to protect this device against high moisture, direct sunlight,shock, solvent, alcohol and gasoline.
- 2. Remove the batteries if the device is being stored for a long time, and keep the batteries far away from children.
- 3. Keep the cuff away from sharp objects and don't extend or twist the cuff.
- 4. This device is not washable. Never immerse the device in water and do not rinse it under the tap. Use only soft and dry cloth to clean the device.
- 5. Do not serve or maintain the cuff and the device when in use with patient.
- 6. The cuff is sensitive and must be handled with care. You can clean the cuff with damp cloth for daily maintenance.To avoid cross infection when sharing the cuff, you can sterilize the fabric cover ofthe cuff with tampons moistened by 3% solution of hydrogen dioxide. After long use there will be a partial discoloration on the fabric surface of the cuff. Do not laundry the cuff as well as ironing with a hot flatiron.
- WARNING:** Under no circumstances may you wash the inner bladder!
- 7. Since neither the device nor batteries are household waste, follow your local recycling rules and dispose them at an appropriate collection site.
- 8. Do not open the device, or delicate electrical components as an intricate air unit could be damaged. If you can not fix the problem using the troubleshooting instruction, please request service from your dealer.
- WARNING:** Do not repair the device without manufacturer's authorization.
- Do not carry out maintenance when using the device.

Caution:

Generally, we recommend the device should be inspected every 2 years and utilize the manometer mode to verify the accuracy of the manometer at least at 50mmHg and 200mmHg after maintenance and repair. Please contact your dealer for maintenance.

SPECIFICATIONS

Model	LD-582
Size	158(L)x120(W)x127(H)mm
Weight	Approximately 490g without batteries
Measuring method	Oscillometry
Extreme Pressure/ cuff pressure	290mmHg
Measuring range	40 to 180mmHg(DIA,diastolic pressure) 60 to 260mmHg(SYS,systolic pressure) 40 to 160 beats/minute(PUL,pulse rate)
Measuring accuracy	± 3 mmHg for static pressure ± 5% of the reading for the pulse rate
Inflation	Automatic by the pump
Rapid deflation	Automatic electronic valve
Batteries	Optional component, 4"AA"x1.5V
Adapter	Optional component,6V.600mA
Memory	2 Users with 60 sets of memory each
Operation temperature and humidity, air pressure	+10°C to+ 40°C,85% and below 800hPa to 1060hPa
Transport and storage temperature and humidity, air pressure	-20°C to +50°C,85% and below 500hPa to 1060hPa
Upper arm circumference	Applicable for arm circumference 22-32 cm (standard cuff)
Complete kit	Main body, storage case, cuff, 4 x AA batteries (Optional), 1 x CR 2025 button battery, adapter (Optional), instruction manual
Pollution Degrees	Degrees 2
Over voltage category	Category II
High Altitudes(m)	≤2000m
Fuse	1A6V 2.1mm*1.45mm*0.81mm

MANUFACTURER'S DECLARATION

Compliance information for each EMC test

Electromagnetic Emission (Home Healthcare Environment)	
Emission test (IEC60601-1-2:2014)	Compliance
Conducted and radiated RF emissions	CISPR 11 Group1 Class B
Harmonic emissions IEC 61000-3-2	Class A
Voltage fluctuations/flicker emissions IEC 61000-3-3	Complies

Compliance information for each EMC test

Declaration -Electromagnetic Immunity (Home Healthcare Environment)		
Immunity test	IEC 60601 test level	Compliance level
Conducted RFIEC 61000-4-6	3V 150 kHz to 80 MHz 6V in ISM and amateur radio bands between 0.15 MHz and 80 MHz	3V 150 kHz to 80 MHz 6V in ISM and amateur radio bands between 0.15 MHz and 80 MHz
Radiated RFEC61000-4-3	10 V/m 80 MHz to 2.7GHz also meet the requirement of table 9 of IEC 60601-1-2:2014.	10 V/m 80 MHz to 2.7GHz also meet the requirement of table 9 of IEC 60601-1-2:2014.
Electrostatic discharge (ESD) EEC 61000-4-2	+8kV contact +2 kV, +4 kV, ±8 kV, +15 kV air	+8kV contact±2 kV, ±4 kV, ±8 kV, ±15 kV air
Electrical fast transient/burst IEC 61000-4-4	+2kV for power supply lines	+2kV for power supply lines
Surge IEC 61000-4-5	+0.5kV,+ 1 kV line(s) to lines	+ 0.5kV, ± 1 kV line(s) to lines
Voltage dips,short interruptions and voltage variations on power supply input lines IEC 61000-4-11	0% Ut,0.5 Cycle at 0°, 45°, 90° 135°, 180°225°, 270° and 315° 0% Ut,1 Cycle and 70% Ut 25/30 cycles sigle phase: at 0° 0% Ut, 250/300 cycles	0% Ut,0.5 Cycle at 0°, 45°, 90° 135°, 180°225°, 270° and 315° 0% Ut,1 Cycle and 70% Ut 25 cycles sigle phase: at 0° 0% Ut, 250 cycles
Power frequency (50/60Hz) magnetic field IEC 61000-4-8	30 A/m	30 A/m

NOTE: The EUT is the a.c. mains voltage prior to application of the test level.The following phenomenon is still fulfill the requirement of basic safety and essential performance.

\* UT:230V~/50Hz, The pressure of the EUT is deviation the normal value but the value is still more than 10psi when flow is 4.5l/min.

\*\*UT:230V~/50Hz, The EUT stop working when adding 0% UT, but the EUT can restore its normal mode automatically.

- Use of this equipment adjacent to or stacked with other equipment should be avoided because it could result in improper operation.If such use is necessary this equipment and the other equipment should be observed to verify that they are operating normally.
- Portable RF communications equipment(including peripherals such as antenna cables and external antennas) should be used no closer than 30cm(12 inches to any part of this device, including cables specified by the manufacturer. Otherwise,degradation of the performance of this equipment could result.
- Under the test condition specified in immunity, the product can provide the basic safety and essential performance.If the essential performance is lost or degraded, additional measures are necessary,such as reorienting or relocating the device.

QUALITY GUARANTEE

MODEL		
Warranty Period	Two-years from purchasing date	
Purchasing date		
Purchasing shop	Name:	Telephone:
	Address:	
Customer	Name:	Telephone:
	Address:	

- 1. Warranty for this automatic digital blood pressure monitor is 24 months from the date of purchase. The 24 months warranty excludes the monitor cuff. The cuff is warranted for 12 months.
- 2. The warranty obligations are prescribed by by warranty certificate for buyer.
- 3. The addresses of organizations for guarantee maintenance are present in the warranty certificate.

WARNING

Do not modify this equipment without authorization of the manufacturer.

All major maintains on the device must be performed by an authorized service center or distributor. No use-serviceable parts inside, before servicing to authorized representative or manufacturer!

DECLARATION:

When technical information for user or service personnel requirements is not in the scope of confidentiality of the Company, the Company committed to provide information disclosure in accordance with procedure, including circuit diagrams and parts lists, and other related type technology information that do not involve commercial secrets may be disclosed, Access to information channels and procedures,please contact your dealer or manufacturer.

REQUIRING RECORD

Date	TROUBLE	SERVICE MAN
Guarantee Regulation	1. During warranty period the repair could be made at any BPM repair department. 2. The following things not belong to warranty range: (1) Operating BPM different from procedures or instructions of the manual. (2) The body is damaged artificially. (3) Self-repairing or modifying the monitor construction in any way. (4) Breakdown due to corrosion of battery leakage. (5) Problem which occurs under natural calamity and other force maieures.	

PERIODIC SAFETY CHECKS

If you use the device with power adapter, preventive inspection and maintenance to be performed including the frequency of such maintenance.

Every time before use, please check the adapter, once damaged, never to use.

Please clean the plug of adapter plug at least once a year. Too much dust on plug may cause the fire. The manufacturer reserves the right to make technical changes without notice in the interest of progress.Prior notices will not be given in case of any amendments within this manual. The mentioned trademarks and names are owned by the corresponding companies.

GIMA WARRANTY TERMS

The Gima 12-month standard B2B warranty applies.

**Disposal:** The product must not be disposed of along with other domestic waste. The users must dispose of this equipment by bringing it to a specific recycling point for electric and electronic equipment.



**HONSUN (NANTONG) Co., Ltd.**

Address: No.8, Tongxing Road, Economic&Technical  
Development Area, Nantong City, Jiangsu, P.R.China  
Made in China



**SHANGHAI INTERNATIONAL HOLDING CORP.GMBH (EUROPE)**

Address: Eiffestrasse 80, 20537 Hamburg GERMANY



LD-582 (GIMA 32778)



Gima S.p.A. Via Marconi,1 20060 Gessate (Mi) - Italy  
gima@gimaitaly.com - export@gimaitaly.com  
www.gimaitaly.com



P582/1901/05

M32778-M-Rev007.24