

Alltion (Guangxi) Instrument Co., Ltd.

EU DECLARATION OF CONFORMITY

Manufacturer name and address:

Alltion (Guangxi) Instrument Co., Ltd.

No. 10, 3rd Road, Wuzhou Industrial Park, Wuzhou City,

Guangxi Province, China. Tel: + 86-774-2836101

Fax: + 86-774-2836192 Postal Code: 543000 mail: sales@alltion.com

SRN (Single Registration

Number):

CN-MF-000024670

Authorized Representative name and address (if applicable):

Obelis s.a.

Bd. Général Wahis 53 B-1030 Brussels, Belgium Phone: 32.2.732.59.54 Fax: 32.2.732.60.03 E-mail: mail@obelis.net

SRN (Single Registration Number):

BE-AR-000000106

We, Alltion (Guangxi) Instrument Co., Ltd. hereby declare under our sole responsibility that the devices listed in the attached Annex comply with the provisions of the REGULATION (EU) 2017/745.

The devices are Class I following Rule 13 of Annex VIII of REGULATION (EU) 2017/745.

The Conformity Assessment Procedure has been performed following MDR Article 19 and Article 52(7), and Technical Documentation was drawn up following Annex II and III of the Regulation (EU) 2017/745.

The following Standards have been used and in relation to which conformity is declared: IEC 60601-1:2005 + CORR.1:2006 + CORR.2:2007 +A1:2012

EN 60601-1:2006+A1:2013

IEC 60601-1-2:2014

EN 60601-1-2:2015

ISO 14971:2019

ISO 15223-1:2021

The present declaration is in conformity with Regulation (EU) 2017/745.

PRRC Signature:

Yuangang Jiang

Position: Vice General Manager

Place of Issue: Guangxi, China

Date of Issue: Jun 30,2022



Annex - List of Devices Covered by the Declaration of Conformity

#	Commercial name/ Trade name	Model/ Reorder/ Catalogue no.#	Description	Intended Purpose	Class	Classificatio n rule applied	BASIC UDI - DI	GMDN/ EMDN code
1	AC-1000 Colposcope	AC-1110 AC-1310 AC-1320	Be used for magnification and illumination of gynecological microsurgery	The ALLTION brand Colposcope is a device designed to permit viewing of the tissues of the vagina and cervix by a telescopic system located outside the vagina. A colposcope is used to diagnose and examine abnormalities of the vagina and cervix.	I	Rule 13	697044351 AC1000XP	10960