Tel: +886-2-6625-8188 Fax: +886-2-6625-0288

EC Declaration of Conformity

We, TaiDoc Technology Corporation

www.taidoc.com

B1-7F, No.127, Wugong 2nd Road, Wugu Dist., 24888 New Taipei City, TAIWAN declare under our sole responsibility that the product

Product Name

: Hemoglobin test strip

Product Model

: TD-4673

Classification

: 98/79/EC (IVDD), Annex II, Self-testing

Conformity Assessment Route : 98/79/EC (IVD), Annex IV excluding section 4&6

EC Certificate Number

: V1 052126 0069 Rev.03

European Representative

: MedNet EC-REP GmbH

Borkstraße 10, 48163 Münster, Germany

Notified Body (CE0123)

: TÜV SÜD Product Service GmbH

Ridlerstraße 65, 80339 München, Germany

GMDNS code

: 57273

to which this declaration relates is in conformity with the following standard(s) or other normative document(s):

Standard	Title
EN ISO 14971:2012	Medical devices - Application of risk management to
	medical devices
EN ISO 23640:2015	In vitro diagnostic medical devices. Evaluation of
	stability of in vitro diagnostic reagents
EN ISO 18113-4:2011	In vitro diagnostic medical Devices. Information
	supplied by the manufacturer (labelling)In vitro
	diagnostic reagents for self-testing
	In vitro diagnostic medical Devices. Information
EN ISO 18113-2:2011	supplied by the manufacturer (labelling). In vitro
	diagnostic reagents for professional use
	Medical devices. Symbols to be used with medical
EN ISO 15223-1:2021	device labels, labelling and information to be
	supplied. General requirements
EN ISO 13485:2016	Medical devices - Quality management systems -
	Requirements for regulatory purposes



TaiDoc Technology Corp. | 81-7F., No.127, Wugong 2110 Ro., New Taipei City 24888, Taiwan

泰博科技股份有限公司 | 新北市24888五股區五工二路127號81-7樓 B1-7F., No.127, Wugong 2nd Rd., Wugu Dist.,

Tel:+886-2-6625-8188 Fax:+886-2-6625-0288

www.taidoc.com	
EN 13532:2002	General requirements for in vitro diagnostic medical
	devices for self-testing
EN 13612:2002/AC:2002	Performance evaluation of in vitro diagnostic medical
	devices
EN ISO 17511:2003	In vitro diagnostic medical devices - Measurement of
	quantities in biological samples - Metrological
	traceability of values assigned to calibrators and
	control materials (ISO 17511:2003)
EN ISO 18113-1:2011	In vitro diagnostic medical devices - Information
	supplied by the manufacturer (labelling) - Part 1:
	Terms, definitions and general requirements

2023, 1,5

Date of Issue

Management Representative Jim Jan