

EUROPEAN MEDICAL DEVICE REGULATION

Declaration of Conformity

As Legal Manufacturer, we

3M Company Single Registration Number: US-MF-000014086 2510 Conway Ave. St. Paul, MN 55144 USA

hereby declare under our sole responsibility that the following CE marked device(s)

	The state of the s
Trade Name	3M [™] Curos Tips [™] Disinfecting Cap Strip for Male Luers
Intended Purpose	The Curos TM Disinfecting Cap for Needleless Connectors, as a medical
	device, is intended for use on needleless connectors only as a disinfecting
	cleaner prior to I.V. access and to act as a cover between line accesses. The
	cap will disinfect the needleless connector one (1) minute after application
	and protect from contamination between accesses for up to seven (7) days
	if not removed. The effectiveness of the cap was tested in vitro against
	Staphylococcus aureus, Staphylococcus epidermidis, Escherichia coli,
	Pseudomonas aeruginosa, Candida glabrata, and Candida albicans and
	was found to have >4 log reduction. The cap may be used in the home or
	healthcare facility.
Reference	CM5-200R
Basic UDI-DI	06082238401010000000075AR

are classified per rule 16 of Annex VIII of the Medical Device Regulation (EU) 2017/745, as Class IIa devices in accordance with Annex XI Part A and all other applicable provisions of the REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL concerning medical devices.

This declaration is made based on the quality assurance certificate EC Certificate Number: MDR 725202

Issued by: BSI, 2797



The Authorized European Representative for the concerned device(s) is

3M Deutschland GmbH Health Care Business

Single Registration Number: DE-AR-000011642

Carl-Schurz-Str. 1

41453 Neuss, Germany

Dianne Gibbs, RAC

Regulatory Affairs Director

3M Medical Solutions Division, 3M Company

2510 Conway Ave. St. Paul, MN 55144 USA

3M is a trademark of 3M.