

EUROPEAN MEDICAL DEVICE REGULATION

Declaration of Conformity

As Legal Manufacturer, we

3M Company Single Registration Number: US-MF-000014086 2510 Conway Ave. St. Paul, MN 55144 USA

hereby declare under our sole responsibility that the following CE marked device(s)

Trade Name	3M TM Curos TM Disinfecting Cap for Needleless Connectors
	3M TM Curos TM Disinfecting Cap Strip for Needleless Connectors
	3M TM Curos TM Disinfecting Cap for Tego® Hemodialysis Connectors
	3M TM Curos Jet TM Disinfecting Cap Strip for Needleless Connectors
	3M TM Curos Jet TM Disinfecting Cap for Needleless Connectors
Intended Purpose	Disinfecting cap for use on needleless connectors only prior to I.V. access
	and to act as a cover between line accesses. The cap will disinfect the
	needleless connector one (1) minute after application and protect from
	contamination between accesses for up to seven (7) days if not removed.
Reference	CFF1-270R: Individual caps, 270 caps per carton
	CFF10-250R: 10-cap strips, 250 caps per carton
	CTG1-270R: Individual caps, 270 caps per carton
	CFJ5-250R: 5-cap strips, 250 caps per carton
	CFJ1-270R: Individual caps, 270 caps per carton
Basic UDI-DI	0608223840101000000040A6

are classified per rule 16 of Annex VIII of the Medical Device Regulation (EU) 2017/745, as Class IIa devices in accordance with Annex XI Part A and all other applicable provisions of the REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL concerning medical devices.

This declaration is made based on the quality assurance certificate EC Certificate Number: MDR 725202 Issued by: BSI, 2797

EU Authorized Representative:

3M Deutschland GmbH Health Care Business Single Registration Number: DE-AR-000011642 Carl-Schurz-Str. 1 41453 Neuss, Germany



— DocuSigned by: Dianne Gibbs

St. Paul 8/29/2022

Location/Date

Dianne Gibbs, RAC Regulatory Affairs Director 3M Medical Solutions Division, 3M Company 2510 Conway Ave. St. Paul, MN 55144 USA

3M is a trademark of 3M.