

EU DECLARATION OF CONFORMITY

HEINE Optotechnik GmbH & Co. KG Dornierstr. 6, 82205 Gilching, Germany www.heine.com

Single Registration Number: DE-MF-000006269

INDIRECT OPHTHALMOSCOPES

We hereby declare, under sole responsibility, that the devices covered by the present EU declaration of conformity are in accordance with the MDR 2017/745.

The indirect ophthalmoscopes consist of following combinations of

| Product name | Basic UDI-DI | GMDN | UMDNS |
|-----------------------------------|----------------------|-------|--------|
| OMEGA 600 | 4053755_IO_01_5Z | 46790 | 12-818 |
| OMEGA 600 wired | | | |
| OMEGA 500 LED | | | |
| OMEGA 500 | | | |
| SIGMA 250 | | | |
| SIGMA 250 M2 | | | |
| Indirect Ophthalmoscope MONOCULAR | - 4053755_HHIO_01_Q9 | 46788 | 12-818 |
| Indirect Ophthalmoscope BINOCULAR | | 46790 | |

| anu | | | |
|------------------------------------|------------------|-------|--------|
| mPack | 4053755_PP_01_97 | | 16-337 |
| mPack UNPLUGGED | | | |
| mPack mini | | | |
| BETA battery handle | 4053755_BH_01_YN | | |
| BETA SLIM battery handle | | | |
| Large battery handle | | | |
| BETA 4 USB rechargeable handle | 4053755_RH_01_79 | | |
| BETA 4 NT rechargeable handle | | | |
| BETA 4 SLIM NT rechargeable handle | | | |
| BETA NT rechargeable handle | | | |
| BETA L rechargeable handle | | | |
| BETA SLIM NT rechargeable handle | | | |
| EN 200 wall transformer | 4053755_WT_01_DE | 16933 | 16-935 |
| EN 200-1 wall transformer | | | |
| | | | |

The indirect ophthalmoscopes are class I according to the risk classification of Annex VIII.

CE

and

References to any common specifications: N/A This declaration of conformity is valid until a revised declaration of conformity is issued.

Gilching, 20 January 2022 (Place and date of issue) Thomas Sauerer / PRRC (Name/function and signature) HEINE OPTOTECHNax GmbH & Co. KG Dornierstr. 6 82205 Gilching