



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-244.10.08



Product Service

EC Certificate

Full Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4)

(Devices in Class IIa, IIb or III)

No. G1 073322 0018 Rev. 01

Manufacturer:

Beijing Target Medical Technologies Inc.

No.60, Shunren Rd.

Shunyi District

101300 Beijing

PEOPLE'S REPUBLIC OF CHINA

Product Category(ies): Disposable Introducer Kit, Disposable Central Venous Catheter Kit, Disposable Blood Pressure Transducer, Arterial Blood Sampling Kits.

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for design, manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex II. This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of class III devices an additional Annex II (4) certificate is mandatory. See also notes overleaf.

Report No.:

BJ19071103

Valid from:

2020-01-03

Valid until:

2024-05-14

Date,

2020-01-03

Christoph Dicks

Head of Certification/Notified Body



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-244.10.08



Product Service

EC Certificate

Full Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4)
(Devices in Class IIa, IIb or III)

No. G1 073322 0018 Rev. 01

Facility(ies):

Beijing Target Medical Technologies Inc.
No.60, Shunren Rd., Shunyi District, 101300 Beijing, PEOPLE'S
REPUBLIC OF CHINA

ZERTIFIKAT ♦ CERTIFICATE ♦ 認證證書 ♦ CERTIFICADO ♦ CERTIFICAT



Product Service

Add value.
Inspire trust.

TÜV SÜD Product Service GmbH · Ridlerstrasse 65 · 80339 Munich · Germany

Beijing Target Medical Technologies Inc.
No.60, Shunren Rd.
Shunyi District
101300 BEIJING
PEOPLE'S REPUBLIC OF CHINA

Your reference/letter of	Our reference/name	Tel. extension/Email	Fax extension	Date	Page
73322	BJ23071100-CL	+86-10-6590-6186 jinglin.chen@tuvsud.com	+86-10-6590-6182	2023-11-02	1 of 4

TÜV SÜD Product Service GmbH Confirmation Letter

CL 073322 0020 Rev. 00

Reference: BJ23071100-CL | 713271022 | 713306516

To whom it may concern,

Confirmation of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 (in the following referenced as MDR) as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices.

With this letter TÜV SÜD Product Service GmbH, designated under MDR and identified by the number 0123 on NANDO, confirms that we have received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the above stated manufacturer with the following SRN Number:

Beijing Target Medical Technologies Inc.
SRN Number: CN-MF-000022386

The devices covered by the formal application and the written agreement mentioned above are identified in the Tables below.

- Table 1 identifies the devices for which an MDR application has been received, written agreement concluded and for which TÜV SÜD Product Service GmbH is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive.
- Table 2 identifies the devices for which an MDR application has been received and a written agreement concluded, but TÜV SÜD Product Service GmbH has not yet taken the responsibility for appropriate surveillance of the corresponding devices under the applicable Directive.

Registered Office: Munich
Trade Register Munich HRB 85742
UniCredit Bank AG · BIC HYVEDEMMXXX
IBAN DE13 7002 0270 0048 8522 11
VAT ID No. DE129484267
Information pursuant to § 2 [1] DL-InfoV
(Germany) at www.tuvsud.com/imprint

Supervisory Board:
Holger Lindner (Chairman)
Board of Management:
Walter Reithmaier (CEO)
Patrick van Welij

Phone: +49 89 50084-747
www.tuvsud.com/ps



TÜV SÜD Product Service GmbH
Munich Branch
Certification Body for Medical Products
Ridlerstrasse 65
80339 Munich
Germany

If devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired after 26 May 2021 and before 20 March 2023, without having been withdrawn, this letter also confirms that

- the manufacturer signed the written agreement under MDR by the date of MDD/AIMDD certificate expiry; or
- provided evidence that a competent authority of a Member State had granted a derogation or exemption from the applicable conformity assessment procedure in accordance with Article 59(1) of MDR or Article 97(1) of the MDR respectively.

The transition timelines in accordance Article 120 (3) of MDR that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120 (3c) of MDR, are shown below:

- 26 May 2026 for Class III custom-made implantable devices
- 31 December 2027 for Class III devices and Class IIb implantable devices (except sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- 31 December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition, measuring function
- 31 December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

The issuance of the first confirmation letter is free of charge. We reserve the right to invoice further copies, amendments and / or changes of the confirmation letter according to effort.

For confirmation letter validity see www.tuvsud.com/ps-cert?q=cert:CL_073322_0020_Rev._00

On behalf of the Notified Body TÜV SÜD Product Service GmbH,
02.11.2023

TÜV SÜD Product Service GmbH
Medical and Health Services



Jinglin Chen
Conformity Assessment Responsible (CARE)

TÜV SÜD Product Service GmbH
Medical and Health Services



Mira Fischer
Application Reviewer

Table 1: Devices covered by this letter and for which TÜV SÜD Product Service GmbH is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive:

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified during application review)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Device 1 Disposable Central Venous Catheter Kit Basic UDI-DI: 69332722MM0017D9	<input checked="" type="checkbox"/> Class III <input type="checkbox"/> Class IIb implantable (non-exempted) <input type="checkbox"/> Class IIb / Class IIb implantable (exempted) <input type="checkbox"/> Class IIa <input type="checkbox"/> Class I devices in sterile condition <input type="checkbox"/> Class I devices with measuring function <input type="checkbox"/> Class III implantable custom-made-device	<input type="checkbox"/> N/A or <input checked="" type="checkbox"/> Identification of the corresponding device under MDD/AIMDD Individual Article number: MMCVCBJ1-14-15; MMCVCBJ1-14-20; MMCVCBJ1-14-30; MMCVCBJ1-16-15; MMCVCBJ1-16-20; MMCVCBJ1-16-30; MMCVCBJ2-70-15; MMCVCBJ2-70-20; MMCVCBJ2-70-30; MMCVCBJ2-70-50; MMCVCBJ3-70-15; MMCVCBJ3-70-20; MMCVCBJ3-70-30	<input checked="" type="checkbox"/> Certification as follows: Certificate G7 073322 0014 Rev.01; NB 0123 Certificate G1 073322 0018 Rev.01; NB 0123 or <input type="checkbox"/> Evidence that a competent authority of a Member State had granted acc. MDR, Art.59 (1) or Art.97 (1) Evidence #1; CA# Evidence #2; CA#
Device 2 Disposable Blood Pressure Transducer Basic UDI-DI: 69332722MM0018DB	<input type="checkbox"/> Class III <input type="checkbox"/> Class IIb implantable (non-exempted) <input checked="" type="checkbox"/> Class IIb / Class IIb implantable (exempted) <input type="checkbox"/> Class IIa <input type="checkbox"/> Class I devices in sterile condition <input type="checkbox"/> Class I devices with measuring function <input type="checkbox"/> Class III implantable custom-made-device	<input checked="" type="checkbox"/> N/A or <input type="checkbox"/> Identification of the corresponding device under MDD/AIMDD Individual Article number:	<input checked="" type="checkbox"/> Certification as follows: Certificate G1 073322 0018 Rev.01; NB 0123 or <input type="checkbox"/> Evidence that a competent authority of a Member State had granted acc. MDR, Art.59 (1) or Art.97 (1) Evidence #1; CA# Evidence #2; CA#

Table 2: Devices covered by this letter and for which TÜV SÜD Product Service GmbH is NOT responsible for appropriate surveillance of the corresponding devices under the applicable Directive:

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified during application review)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Not applicable	<input checked="" type="checkbox"/> N/A	<input checked="" type="checkbox"/> N/A	<input checked="" type="checkbox"/> N/A

Confirmation Letter Version History

Date	TÜV SÜD Product Service GmbH internal reference traceable to each version of the letter	Action
2023/11/02	BJ23071100-CL 713271022 713306516	Initial issuance