3M Center 2510 Conway Ave, Bldg. 275-5W-06 St. Paul, MN 55144 U.S.A. 651 733 1110



Declaration of Conformity

As Legal Manufacturer We,

3M Company 3M Health Care 2510 Conway Ave Saint Paul, MN 55144 USA

hereby declare under our sole responsibility that the CE marked products to which this declaration relates,

3M[™] Curos[™] Disinfecting Cap for Needleless Connectors 3M[™] Curos[™] Disinfecting Cap Strip for Needleless Connectors 3M[™] Curos[™] Disinfecting Cap for Tego® Hemodialysis Connectors

> Product Numbers CFF1-270R, CFF10-250R, CTG1-270R

is classified,
per Rule 15 of Annex IX of the Medical Device Directive 93/42/EEC, as amended per 2007/47/EC
as a Class IIa device
and

is in accordance with Annex V of Directive 93/42/EEC, as amended per 2007/47/EC on the approximation of the laws of the European Union Member States concerning medical devices.

In addition, we declare that the above mentioned devices fulfill the applicable provisions of the Directive 93/42/EEC, as amended per 2007/47/EC.

This declaration is made on the basis of the quality assurance certificate, CE 00493 delivered by BSI, 2797

EU Representative Address 3M Deutschland GmbH Health Care Business Carl-Schurz-Str. 1 41453 Neuss, Germany

____ Date: 21 May 2020

Signature:

Dianne Gibbs

3M Health Care Division Regulatory Affairs Manager

Medical Solutions Division

Page 1 of 1 Pages

REG-DOC-05-293347, Curos Disinfecting Caps Declaration of Conformity, Version 7, Status: Release, Release Date: 05/15/2020 08:40:52 AM CDT