

EU DECLARATION OF CONFORMITY

Name and address of the manufacturer: **Shenzhen Viatom Technology Co., Ltd.
4E, Building 3, Tingwei Industrial Park,
No.6 Liufang Road, Block 67, Xin'an Street,
Baoan District, 518101 Shenzhen, P.R.China**

SRN (Manufacturer): **CN-MF-000012182**

Name and address of Authorized Representative: **MedNet EC-REP GmbH
Borkstrasse 10 , 48163 Muenster, Germany**

SRN (EU Authorised): **DE-AR-000000002**

We declare that the product concerned has been designed and manufactured under a quality management system according to Annex IX of EU 2017/745 (MDR).

Medical Device: **Blood Pressure Monitor
Model: BP2, BP2B, BP2V, BP2W**

Intended use/purpose: **The Blood pressure monitor is intended to record, store, display and transfer single-channel electrocardiogram (ECG), blood pressure and pulse rate in adult population. The device does no analysis by itself and is intended to be used with a compatible ambulatory ECG (Holter) analysis system(AI-ECG Tracker) which will analyze the recorded data (used under the care of a physician). The device data and the data analysis are then reviewed by a trained medical personnel for the purpose of forming a clinical diagnosis.
The device is intended for use by adults' health-conscious individuals.
The device does not include analysis and diagnosis functions.
The device has not been tested and it is not intended for pediatric use.**

GMDN: **45617 Automatic-inflation electronic sphygmomanometer, portable, arm/wrist**

Risk class: **Class IIa**

Basic UDI-DI: **69344401BP2XD**

Conformity assessment procedure: **EU 2017/745 (MDR) Annex IX (Chapter I + III and Sec.4)**

The EU declaration of conformity is issued under sole responsibility of the manufacturer. We hereby declare that the above mentioned products meet the provisions of the following EUROPEAN PARLIAMENT AND OF THE COUNCIL Regulation and Applicable standards. All supporting documents are retained under the premises of the manufacturer.

Regulations:

EU 2017/745 (MDR)
RED, 2014/53/EU
ROHS, (EU) 2015/863
ROHS, Directive 2011/65/EU

Applicable CS or Standard(s):

EN 60601-1:2006/A2:2021
EN 60601-1-2:2015+A1:2021
EN 60601-1-6:2010+A2:2021
EN 60601-1-11:2015/A1:2021
EN 60601-2-47: 2015
EN ISO 10993-1:2020 **EN ISO 10993-5:2009**
EN ISO 10993-10:2023 **IEC 80601-2-30:2018**
EN 62479:2010 **EN 50663:2017**
ETSI EN 300 328 V2.2.2(2019-07)
ETSI EN 301 489-1 V2.2.3 (2019-11)
ETSI EN 301 489-17 V3.2.4 (2020-09)
EN ISO 14971:2019/A11:2021
EN ISO 13485:2016 **EN ISO15223-1: 2021**
EN ISO 20417:2021 **EN 62304:2006+A1:2015**

Certificate No.:

HZ 2120274-1

Issue date:

2024-01-18

Expiry date:

2029-01-17

Notified Body:

TÜV Rheinland LGA Products GmbH
Tillystraße 2
90431 Nürnberg
Deutschland
CE 0197

Shenzhen, 2024/03/15

Place, date



Zhou Saixin General manager

Name and function